FUJIFILM
FUJIFILM Wako Chemicals U.S.A. Corporation

LAL Division
1600 Bellwood Road
Richmond, VA 23237 USA
Phone: (804) 271-7677
Fax: (804) 271-7791

Wako

Certificate of Analysis

Endotoxin Indicator Vial

Product Code:	EIV-0001 (Vial) / EIV-025 (Kit)
Product Description:	Endotoxin Indicator Vial
Lot Number:	EV021901
Production Date:	27 MAR 2019
Expiration Date:	2021.02

Assay:	**Specification:**	**Result:**
Endotoxin Potency (EU per Vial)	>1,000 EU/vial	Pass

Reference
United States Pharmacopeia (USP), 2012, General Chapter <85>, Bacterial Endotoxins Test (BET)

This is to certify that this product lot was manufactured according to current cGMP Standards, meets FUJFILM Wako Chemicals U.S.A. Corporation – LAL Division's product specifications, and complies with all applicable requirements for product release.

Created By: Jason Holliday Date: 07/31/2019

Quality Assurance Approval By: [signature] Date: 7/31/2019