

Certificate of Analysis**Gel Clot Tubes**

Product Code: CT-1075

Product Description: Depyrogenated Lysate Gel Clot Tubes (10 x 75 mm)

Lot Number: 041911

Production Date: 17 APR 2019

Quality Assurance
Release Date: 18 JUN 2019

Expiration Date: 2022.04


Depyrogenation:

Temperature: 255°C ± 5°C Time: 240 minutes

Assay:**Result:**

Bacterial Endotoxin Test <0.001 EU/tube
Interfering Test Pass

This is to certify that this product lot was manufactured according to current cGMP Standards, meets FUJIFILM Wako Chemicals U.S.A. Corporation – LAL Division's product specifications, and complies with all applicable requirements for product release.

Created By:  Date: 06/18/2019Quality Assurance
Approval By:  Date: 06/18/2019